

## Certificate of Compliance

**Issue Date:** 23<sup>rd</sup> October 2018

**Report No.:** 18E7380-1 Issued October 2018  
18S7515-1 Issued October 2018

This is to certify that the following designated product

---

**Product:** NV1050 Airborne Infection Control Unit

**Trade name:** Novaerus

**Model Number:** NV1050

**Company Name:** Novaerus Ltd.

This product, which has been issued with the test report listed as above by Compliance Engineering Ireland Ltd is confirmed to comply with the requirements of the following standards.

### **Standards**

#### **EMC:**

- EN 55011: 2009 + A1: 2010
- EN 60601-1-2: 2014 (4<sup>th</sup> Edition)
- EN 61000-3-2: 2014
- EN 61000-3-3: 2013

#### **Safety:**

- EN 60601-1: 2006 + A1:2013 + A11: 2011

  
John McAuley  
Managing Director